

For Immediate Release: 1 July 2015

**INTERTEK WARNS OF COUNTERFEIT cETLus CERTIFICATION MARK
ON QEL CO GAS DETECTORS**

Intertek Testing Services (ETL) notifies consumers, retailers and distributors that the QEL CO Gas Detectors Models Q5 and B5 may be improperly being used in environments not covered by the product certification as evaluated by Intertek.

The products are intended only for laboratory and industrial applications, not for life safety applications where they are monitoring gas concentrations of occupied spaces.

Please note the following clarification for the proper use of the products:

THESE PRODUCTS ARE NOT INTENDED FOR LIVE-SAFETY APPLICATION NOR EVALUATED FOR SUCH USE. PRODUCT ARE NOT INTENDED TO BE USED AS MEANS TO ACTIVATE VENTILATION SYSTEM IN AN AREA OR USED AS MEANS FOR SIGNALLING AN EVACUATION.

**Name of Product: Gas Monitor
Brand/Manufacturer: QEL, Ltd.
Model: Q5, B5
Sold at: Various Distributors**

Photo of Product:

